

Ref. No. : PUR/30/13/1639-52

Dated : 20.12.2013

Subject : Notice Inviting Quotation for Purchase of Equipment required for Chem. Tech. Deptt.(CAT - A).

This institute intends to purchase of equipment as detailed given below. Interested firm/party are requested to send the quotation to the office of undersigned in a sealed cover superscribed "Quotation for Equipment for Chem. Tech. Deptt." on or before 10.01.2014.

S.No.	Description	Qty.	Remarks
1.	Laboratory Drying Oven.	02 No.	Detailed Specifications as per Annexure - 'A'
2.	Autoclave (Vertical Model).	01 No.	

Note : It may be noted that quotation received only through registered/speed post shall be considered only. The institute is located in remote area and it take 5 to 7 days to reach the mail, therefore, quotation be dispatched well in time considering this factor.

N.B. :

1. Rate of ST/VAT if extra must be mentioned clearly.
2. Price quoted must be FOR SLIET.
3. Quotation received later than due date are liable to be ignored/rejected.
4. Other terms and condition for submitting the quotation are given overleaf which must be read carefully before submitting the quotation.
5. We are not responsible for accidental opening of the cover if it is not properly subscribed and sealed.
6. Quotation must be submitted on letter head of the firm with all particular, any other format will not be acceptable.

**Faculty I/c, Purchase
PTO**

TERMS & CONDITIONS FOR QUOTATION

DELIVERY	The rate quoted must preferably be free delivery/F.O.R. Longowal after allowing the discount, if any. Where quoted extra advalorem rate payable should clearly be indicated. Supply should be made within the specified delivery period.
TERMS OF PAYMENT	Our normal terms of payment is within 30 days after receipt of stores in good condition by means of cheque/draft
TAXES	No sales tax concession against Form C and 'D' is admissible to this Institute. However, form of certificate being an educational institute can be issued if sales tax concession is admissible.
EXEMPTIONS	Excise and customs duties are exempted to the institute. The relevant exemption certificate will be issued to the successful bidder only if the excise duty/custom duty is exclusively mentioned in the Quotation.
DIRECTOR'S RIGHTS	Director, SLIET, reserves the rights of acceptance or rejection of any or all quotations. The discretion for increasing or decreasing of the quantities also rests with him. SLIET also not bind itself to accept does the lowest price. In case of any dispute, the decision of Director SLIET will be final & binding.
VALIDITY OF QUOTATIONS	Quotations will be considered valid for 3 months from the date of receipt.
CORRESPONDENCE	No correspondence regarding acceptance/rejection of a quotation will be entertained.
SAMPLE/BRAND/MAKE/WEIGHT	Sample where asked for, will invariably be made available and sent along with the quotations. However, Brand/Make/Weight etc. must be mentioned clearly in the quotations. Technical literature/pamphlet should also be enclosed.
REJECTION	Quotation not confirming to the set procedure as above will be rejected.
DISCOUNT/REBATES	<p>A special discount/rebate wherever admissible keeping in view that the supplies are being made for education purpose in respect of Public Institution of national importance may please be indicated.</p> <p>Conditional, telegraphic quotation shall be rejected out rightly.</p> <p>SLIET shall not be held responsible for any postal delay in sending or late receipt of quotation.</p> <p>Quotation should be free from corrections & erasures.</p>

Faculty I/c, Purchase

Chemical Technology Department

ANNEXURE - 'A'

S. No.	Name of Equipment	Specifications	Qty.
1.	Laboratory Drying Oven	Interior chamber space of 500mm *500mm *550mm. PID Controller and LED Display with temperature range ambient to 250°C. Readability and set ability of 1°C. Temperature variation of 1°C. Sensor thermocouple type, With fan speed variation (4 steps). 220-250 V, 50 Hz power supply. With exhaust valve and five trays inside at regular interval. Able to maintain uniform temperature inside. With 3 years Warranty.	02
2.	Autoclave (Vertical Model)		01
	Size (Internal Basket):	Width 15 inch, height 25 inch (Minimum)	
	Working Pressure:	15 psi to 20 psi or better	
	Sterilizing Temperature:	121°C to 134°C or better	
	Sterilizing Timer:	0 to 99 Minutes	
	Control:	Microprocessor w/ Digital Display	
	Heat Average:	$\leq \pm 1^\circ\text{C}$	
	Pressure Indicator:	Pressure gauge	
	Construction:	Double / Triple wall construction	
	Inner/outer Chamber:	316 / 316L grade Stainless Steel	
	Insulation:	Glass Wool	
	Waste Container:	Stainless Steel	
	Door / Lid:	Quick door open structure, safety and interlock device	
	Door Locking:	Radial Type	
	Alarm audio/ visual:	Low water level alarm, high temperature alarm	
	load	Should maintain the temperature with in Half hour.	
<p>Safety Features Lever lock for lid, safety valve (emergency pressure release), low water detector and pressure interlock, Temperature-dependent door lock, Excess temperature safety device, Door lock safety switch, Overpressure valve</p>			