



संत लौगोवाल अभियांत्रिकी एवं प्रौद्योगिकी संस्थान,  
लौगोवाल, संगरूर, पंजाब - १४८ १०६

(मानव संसाधन मंत्रालय के अधीन समविश्वविद्यालय)

**Sant Longowal Institute of Engineering and Technology**  
**Longowal, Dist. Sangrur, Punjab - 148106**  
(Deemed to be University under MHRD)

Ref. No. SLIET/PUR/25/19/

Date:

(Registered Post)

M/s \_\_\_\_\_

**Sub: Notice re-inviting quotation for purchase of equipment under CSIR project.**

This Institute intends to purchase equipment under CSIR project. Please send quotation to the undersigned in a sealed cover duly super scribed on envelop "**Quotation for purchase of equipment**" so as to reach latest by closing date i.e. 03.01.2020. The detailed specifications are as under:

S. No.	Detailed Specification	Qty.	EMD	Remarks
1.	Rotary Vacuum Evaporator System (Detailed technical specifications are enclosed at Annexure-A)	01	15000/-	

**The Terms & Conditions for submitting quotation are as under:-**

- 1) It may be noted that quotation received only through registered/speed post shall be considered. The institute is located in a remote area and it takes 5 to 7 days to reach the mail, therefore, quotation be dispatched well in time to avoid any sort of delay.
- 2) Rate of GST/IGST, if extra must be mentioned clearly along with GST number.
- 3) The other terms & conditions for submitting the quotation are given overleaf which must be read carefully before submitting the quotation.
- 4) The Price quoted must be FOR SLIET, Longowal.
- 5) Quotations received later than due date are liable to be rejected and returned as it is.
- 6) Quotation must be sent on the letter head of the party duly addresses to Deputy Registrar (Store & Purchase) along with a certificate that the firm/company is not blacklisted by any govt. organization/dept.
- 7) As per instructions if purpose of quotation is not super scribed and quotation is opened by mistake then it may be rejected.
- 8) The bid security of **Rs. 15,000/-** is required to be submitted in the form of Account Payee Demand Draft, Fixed Deposit Receipt, Banker's Cheque or Bank Guarantee from any of the Commercial Banks in favour of Director, SLIET, Longowal.

Deputy Registrar  
(Store & Purchase)

## TERMS & CONDITIONS FOR QUOTATION

DELIVERY	The rate quoted must be free delivery/F.O.R. Longowal after allowing the discount, if any. Where quoted extra ad-valorem rate payable should clearly be indicated. Supply should be made within the specified delivery period.
TERMS OF PAYMENT	Our normal term of payment is within 45 days after receipt of stores in good condition by means of cheque/draft/RTGS.
PERFORMANCE SECURITY	Performance Security @ 5% of order value in the shape of DD, FDR or BG valid for a period of sixty days beyond the date of completion of all contractual obligations of the supplier including warranty obligations, as per GFR-2017 rules, is also required.
Concessional GST	The Institute is registered with the department of Scientific & Industrial Research (DSIR) for purposes of availing customs duty exemption in terms of Govt. notification No. 51/96-customs dated 23.07.1996 and central excise duty exemption in terms of Govt. notification No. 10/97-central excise dated 01.03.1997 and 45/2017 & 47/2017 dated 14/11/2017 regarding GST as amended from time to time. The relevant exemption certificate will be issued to the successful bidder only if the excise duty/customs duty/GST are exclusively mentioned in the commercial bid of e-Tender/NIQ. <b>NO ADVANCE PAYMENT WILL BE MADE.</b>
EXEMPTIONS	Excise and customs duties are exempted to the institute. The relevant exemption certificate will be issued to the successful bidder only if the excise duty/custom duty is exclusively mentioned in the Quotation. In case the offered items are to be imported, the rates should be quoted in foreign currency on FOB basis. Basic duty as applicable under notification No. 51/96 customs dated 23.07.1996 as applicable shall be borne by the institute. If the price quoted is in foreign currency then payment shall be made through letter of credit against submission of B.G. of Min. 25% value of supply order, if the order value is more than US\$ 10,000 or through Telegraphic Transfer (TT) if the order value is less than US\$ 10,000 through Nationalized banks. The bank charges outside India should be borne by the Beneficiary. Clearance at customs will be arranged by us but you will assist our clearance agent. In case of indigenous item the price must be quoted in Indian Rupees and 100% payment will be made only after successful installation, testing and commissioning of equipment.
DIRECTOR'S RIGHTS	Director, SLIET, reserves the rights of acceptance or rejection of any or all quotations. The discretion for increasing or decreasing of the quantities also rests with him. SLIET also does not bind itself to accept the lowest price. In case of any dispute, the decision of Director SLIET will be final & binding.
VALIDITY OF QUOTATIONS	Quotations must remain valid for period of at least 03 months.
SAMPLE/BRAND/MAKE/WEIGHT	Sample, where asked for, will invariably be made available and sent along with the quotations. However, Brand/Make/Weight etc. must be mentioned clearly in the quotations. Technical literature/pamphlet should also be enclosed.
REJECTION	Quotation not confirming to the set procedure as above will be rejected. Conditional, telegraphic quotation shall be rejected out rightly.
DISCOUNT/REBATES	A special discount/rebate wherever admissible keeping in view that the supply is being made for education purpose in respect of Public Institution of national importance may please be indicated.
GENERAL TERMS	SLIET shall not be held responsible for any postal delay in sending or late receipt of quotation. Quotation should be free from corrections & erasures. <b>Other terms &amp; Conditions will be applicable as per GFR-2017.</b>

**Deputy Registrar  
(Store & Purchase)**

# ANNEXURE - II

## **TECHNICAL SPECIFICATION FOR ROTARY VACUUM EVAPORATOR SYSTEM**

### **ROTARY EVAPORATOR**

- Autolift or Manual quick –action jack lifting for raising and lowering of flask
- Rotation speed up to 280 rpm or better with microprocessor control.
- Heating bath for temperature of room temperature to 180°C
- Digital bath having capacity of 5 Liters or more and resolution of 1° C
- Digital bath temperature display of both set and actual temperature simultaneously
- Over-temperature protection power cut off when actual temperature exceeds set temperature or by rapid temperature increase
- Evaporating flask from 50-2000 mL can be used on the same joint adapter without additional connections.
- Vertical Glass Assembly
- 1 liter Evaporating Flask and 1 liter Receiving Flask should be provided in standard scope of supply.
- Automatic over heat cut-off protection
- CE and RoHS compliant

### **VACUUM PUMP**

- Three stage diaphragm vacuum pump with controller
- Chemical resistant diaphragm made of PTFE should be ideal for various applications involving aggressive solvents & acidic vapors.
- Max. vacuum – 2 mbar or better
- Noise level should be 60db or less.

### **CHILLER**


Compact and Robust Re-circulating Chiller with a cooling capacity of 300 Watts


- Temperature Range: -20 to 100°C.
- Working Range: -10 to 80°C
- Coolant: CFC Free

Warranty should be minimum 02 years or more.

The equipment should be supplied with all the necessary accessories and spareparts for a fully functional unit.

Installation & demonstration should be scope of supplier at our premises.

  
(Dr. H. Chopra)  
4/12/19

  
(Dr. P. S. Panesar)  
4/12/19